



Ref. Certif. No.

HU-002081

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Cycling Ergometers

Name and address of the applicant

LODE B.V.  
Zernikepark 16, 9747 AN Groningen,  
The Netherlands

Name and address of the manufacturer

LODE B.V.  
Zernikepark 16, 9747 AN Groningen,  
The Netherlands

Name and address of the factory

LODE B.V.  
Zernikepark 16, 9747 AN Groningen,  
The Netherlands

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

967900; 967912: 100-240V; 50/60Hz; 160W;  
967902; 967922: 100-240V; 50/60Hz; 160W;  
...  
(See summary table on page 2 for further ratings)

Trademark (if any))



Customer's Testing Facility (CTF) Stage used

---

Model / Type Ref.

967900; 967912 - (Angio Single Set)  
967904; 967924 - (Angio Automatic Stand)  
967905; 967925 - (Angio with fixation set)  
967930; 967940 - (Angio imaging with stress support  
imaging; Nuclear Imaging table with Angio)  
967931; 967941 - (Angio imaging with stress support  
ultrasound; ECST Table with Angio)  
967800 - (Bed Cart for Angio with fixation set)  
967902; 967922 - (Angio Static Wall)  
967903; 967923 - (Angio Electric Wall)

Additional information (if necessary may also be reported on page 2)

---

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005 + COR1:2006 + COR2:2007 + AMD1:2012  
IEC 60601-1-6:2010 + AMD1:2013  
National Differences: CA, US

As shown in the Test Report Ref. No. which forms part of this Certificate

28237317 001

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TÜV Rheinland InterCert Kft., Division MEEI  
H-1132 Budapest, Váci út 48/A-B  
www.tuv.hu



Signature:

*Gabor Kassai*  
Gabor Kassai

Date: 2017-10-18



Ref. Certif. No.

HU-002081

**Summary of different types Angio series ergometers**

Model	Type	Rated Voltage (AC, 50/60Hz)	Rated Input (W)
Angio Single Set	967912, 967900	100 – 240V	160
Angio Static Wall	967902, 967922	100 – 240V	160
Angio Electric Wall	967903, 967923	100 – 240V	250
Angio Automatic Stand	967904, 967924	100 – 240V	250
Angio with Fixation set	967925, 967905	100 – 240V	160
Bed Cart for Angio with Fixation set	967800	n.a.	n.a.
Angio imaging with stress support imaging (Nuclear Imaging table with Angio)	967930	230V	400
Angio imaging with stress support imaging 115V (Nuclear Imaging table with Angio)	967940	115V	400
Angio imaging with stress support ultrasound (ECST Table)	967931	230V	260
Angio imaging with stress support ultrasound 115V (ECST table)	967941	115V	260

**Additional information (if necessary)**

Date: 2017-10-18

TÜV Rheinland InterCert Kft., Division MEE  
 H-1132 Budapest, Váci út 48/A-B  
 www.tuv.hu

Signature:



Gabor Kassai