

# Certificate



Certificate no.

CU 72081030 01

**License Holder:**

Lode B.V. Medical Technology  
Zernikepark 16  
9747 AN Groningen  
Netherlands

**Manufacturing Plant:**

Lode B.V. Medical Technology  
Zernikepark 16  
9747 AN Groningen  
Netherlands

**Test report no.:** USA-DS 30793713 001

**Client Reference:** Jaap Noordmans

**Tested to:** UL 60335-1:2004  
ASTM F1250-00  
CAN/CSA E335-1/3E

**Certified Product:** Ergometer

**License Fee - Units**

Model Designation: Corival Series

7

Rated Voltage: AC 115V, 50/60Hz  
Rated Power: 130VA  
Protection Class: I

Special Remarks: This equipment was also evaluated to  
UL 60601-1:2003 and CAN/CSA C22.2 No. 601-M90.

7

Appendix: 1, 1-2

Licensed Test mark:



Signature

Dipl.-Ing. M. Raap  
QA Certification Officer

Date of Issue  
(day/mo/yr)

07/07/2008